



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/698,131

11/01/2003

Martin T. Gerber

BUG8-44251

2892

116 7590 12/17/2008  
PEARNE & GORDON LLP  
1801 EAST 9TH STREET  
SUITE 1200  
CLEVELAND, OH 44114-3108

EXAMINER

LACYK, JOHN P

ART UNIT

PAPER NUMBER

3735

MAIL DATE

DELIVERY MODE

12/17/2008

PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

<b>Office Action Summary</b>	<b>Application No.</b> 10/698,131	<b>Applicant(s)</b> GERBER ET AL.	
	<b>Examiner</b> John P. Lacyk	<b>Art Unit</b> 3735	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 01 December 2008.
- 2a) ☐ This action is **FINAL**.                      2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-9 and 33-36 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☒ Claim(s) 6-9, 33-36 is/are allowed.
- 6) ☒ Claim(s) 1-5 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \*    c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |  |   |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)                                | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                       | 5) <input type="checkbox"/> Notice of Informal Patent Application                       |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)<br>Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____  |

A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on 12/01/08 has been entered.

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-5 are rejected under 35 U.S.C. 103(a) as being unpatentable over Johnson et al (6,338,345) in view of Goupil (6,652,883).

Johnson et al discloses a device used to deliver a bulking prosthesis to the body.

Johnson et al teaches using an endoscope having a opening (58) or cavity for receiving tissue from the target site, the tissue is drawn into the opening by a vacuum and a needle is used to make a hole or puncture in the tissue and a "pushing agent" to push the prosthesis from the distal end of the tube through the hole. With respect to the bulking prosthesis Johnson et al teaches (column 16, lines 1-21) that it is well known to use a hydrophilic/hydrogel material as the bulking prosthesis. Also Johnson et al teaches (column 6, line 34-column 7, line 18) that the bulking prosthesis can take on a wide variety of shapes and sizes and that optimal dimensions are patient specific and can be determined through routine experimentation of one skilled in the art.

Johnson et al discloses a device for treating gastroesophageal reflux disease (GERD) and discloses the claimed device and method except for specifically teaching using the bulking prosthesis for treating urinary incontinence. Goupil et al teaches that it is well known to use a bulking material to treat a variety of problems including GERD and urinary incontinence. Therefore a modification of Johnson et al such that the bulking prosthesis is used to treat urinary incontinence would have been obvious in view of Goupil et al which shows that it is well known to treat both with a bulking prosthesis.

The drawings are objected to under 37 CFR 1.83(a). The drawings must show every feature of the invention specified in the claims. Therefore, the balloon at the distal end of the tubular instrument, as discussed in paragraph 0037, must be shown or the feature(s) canceled from the claim(s). No new matter should be entered.

Corrected drawing sheets in compliance with 37 CFR 1.121(d) are required in reply to the Office action to avoid abandonment of the application. Any amended replacement drawing sheet should include all of the figures appearing on the immediate prior version of the sheet, even if only one figure is being amended. The figure or figure number of an amended drawing should not be labeled as "amended." If a drawing figure is to be canceled, the appropriate figure must be removed from the replacement sheet, and where necessary, the remaining figures must be renumbered and appropriate changes made to the brief description of the several views of the drawings for consistency. Additional replacement sheets may be necessary to show the renumbering of the remaining figures. Each drawing sheet submitted after the filing date of an

Art Unit: 3735

application must be labeled in the top margin as either "Replacement Sheet" or "New Sheet" pursuant to 37 CFR 1.121(d). If the changes are not accepted by the examiner, the applicant will be notified and informed of any required corrective action in the next Office action. The objection to the drawings will not be held in abeyance.

Applicant's arguments filed 12/01/08 have been fully considered but they are not persuasive. Applicant argues that Goupil teaches injection of a bulking material into the submucosa of the urethra but fails to disclose where the bulking agent should be injected, specifically through a hole proximate to an external urethral sphincter. As pointed out by Applicant Goupil does teach injecting the material into the submucosa of the urethra. Scopelianos et al is cited (column 7, line 58-column 8, line 10) to provide a teaching of a similar bulking material that is injected into the submucosa of the urethra (at or around the urethral-bladder junction to the external sphincter). This provides a teaching that one skilled in the art would know that injecting an agent into the submucosa of the urethra would include being proximate to the external urethral sphincter and that Goupil would in fact teach injecting the bulking material at the desired location.

Claims 6-9 and 33-36 are allowed.

Art Unit: 3735

Any inquiry concerning this communication or earlier communications from the examiner should be directed to John P. Lacyk whose telephone number is (571)272-4728. The examiner can normally be reached on 8-4:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Chuck Marmor, II can be reached on 571-272-4730. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

J.P. Lacyk

/John P Lacyk/  
Primary Examiner, Art Unit 3735